


EnviroSoft[®] Supine Head Positioner Cleaning Protocol


**Avoid alcohol-based or strong, undiluted disinfectants*

1. Using disposable disinfectant wipes, while holding the positioner, wipe evenly on both sides:



2. While holding the positioner, use disposable disinfectant wipes on both the outer diameter and inner diameter.

Outer Diameter



Inner Diameter

